



STATE OF CONNECTICUT

DEPARTMENT OF CONSUMER PROTECTION

Report of Test for: 1 - Set of Avoirdupois Weights 5 pound to 0.01 pound and 1 ounce (13 Weights)

Submitted by:
ADVANCED SCALE, INC.
 13 DELTA DRIVE UNIT 6
 LONDONDERRY, NH 03053

TESTED BY: DEPT. OF CONSUMER PROTECTION, DIVISION OF FOOD AND STANDARDS - METROLOGY LABORATORY
 9 WINDSOR AVENUE, WINDSOR, CT 06095

Manufacturer: Unknown
 S/N: VT05-485

CT. TEST NO. 2019 - 63

The weights described above have been compared with the Standards of the State of Connecticut (Working Standards - Avoirdupois Sets F and G), traceable to the International System of Units (SI), through the Report of Calibration Maine Test No. 7124ME and Rice Lake Traceable Calibration Certificate No. a2412935 and were found to be appropriate for their intended use and to be accurate / (adjusted) to within CLASS F TOLERANCE as indicated on the attached tolerance table.

The following mass values reflect the as found and as left condition of weights as they were delivered for testing.

DENOMINATION	CLASS F TOLERANCE (± MILLIGRAM)	AS FOUND CONVENTIONAL MASS CORRECTION (MILLIGRAM)	AS LEFT CONVENTIONAL MASS CORRECTION (MILLIGRAM)	UNCERTAINTY (± MILLIGRAM)	K COVERAGE FACTOR
5 lb F	230	48	SAME	15	2.02
2 lb F	91	24.0	SAME	5.7	2.05
2 lb * F	91	19.5	SAME	5.7	2.05
1 lb F	70	18.2	SAME	4.4	2.02
0.5 lb F	45	9.3	SAME	2.8	2.02
0.2 lb	18	6.7	SAME	1.1	2.02
0.2 lb *	18	6.7	SAME	1.1	2.02
0.1 lb	9.1	2.81	SAME	0.58	2.02
0.05 lb	4.5	0.07	SAME	0.30	2.02
0.02 lb	1.8	0.32	SAME	0.15	2.02
0.02 lb *	1.8	0.30	SAME	0.15	2.02
0.01 lb	1.5	0.574	SAME	0.095	2.02
1 oz	5.4	1.09	SAME	0.35	2.02

Test Procedure: SOP #8, MODIFIED SUBSTITUTION
 Balances used: PM5003, AT1005, AT106, UMT5 (METTLER TOLEDO balances)
 Temperature: 20.4 °C to 20.7 °C, Humidity: 43 % to 44 %, Pressure: 768.0 mmHg to 769.1 mmHg

DATE RECEIVED: 03/22/2019
 DATE OF TEST: 03/28/2019

Signing Authority

METROLOGIST: ANA MARIA FELICIANO
 FOOD AND STANDARDS DIVISION

- Notes:
- (1) This report relates only to the items listed in this report at the time of test.
 - (2) The uncertainties are calculated according to the Guide to the Expression of Uncertainty in Measurement, where the expanded uncertainty is k times the root sum square of the Type A and Type B uncertainties. The coverage factor (k) defines a level of confidence of approximately 95 %.
 - (3) The effect of air buoyancy has been evaluated and determined to be insignificant.
 - (4) This report shall not be reproduced except in full, without written approval of the laboratory.
 - (5) THE REPORT OF TEST IS INVALID WITHOUT THE OFFICIAL STATE OF CONNECTICUT WEIGHTS AND MEASURES SEAL AFFIXED.