

625 East Bunker Court
 Vernon Hills, Illinois 60061
 PH: 866-466-6225
 Fax: 847-327-2993
 www.innocalsolutions.com

NIST Traceable
Calibration Report

REPORT NUMBER
1700595

Reference Number: 116455700
 PO Number: NBROWN061022

Advanced Scale Inc
 13 Delta Dr
 Ste 6
 Londonderry, NH 03053 United States

Manufacturer:	Traceable	Calibration Date:	06/15/2022
Model Number:	4096	Calibration Due Date:	06/15/2024
Description:	Humidity, Temp Meter 25 to 95%RH 0-50°C	Condition As Found:	In Tolerance
Asset Number:	CP373852	Condition As Left:	In Tolerance, No adjustment
Serial Number:	200074629		
Procedure:	DS Control Company 4096		
Remarks:	NIST-traceable calibration performed on the unit referenced above in accordance with customer requirements, published specifications and the lab's standard operating procedures. No adjustments were made to the unit.		

Standards Used

Standard ID	Manufacturer	Model Number	Description	Cal Date	Due Date
CP144802	Rotronic Instrument Corp	HC2A-S	Humidity, HygroClip2	6/21/2021	6/30/2022

Calibration Data

Function Tested	Nominal / Reference Value	Measured Value	OOT	Calibration Tolerance <i>g: = Guard Banding Applied</i>	TUR	EMU
Relative Humidity	50.0 %RH					
As Found & As Left	50.0	50		46.0 to 54.0 %RH	4.9:1	± 0.82 %RH
Temperature	25.0 °C					
As Found & As Left	25.0	24		24.0 to 26.0 °C	8.5:1	± 0.12 °C

Temperature: 26 °C
Humidity: 65 %RH
Rpt. No.: 1700595

Calibration Performed By:				Quality Reviewer:	
Elaine Yonkus		Metrologist		Szplit, Tony	6/15/2022
<small>Name</small>	<small>ID#</small>	<small>Title</small>	<small>Phone</small>	<small>Name</small>	<small>Date</small>

This report may not be reprinted, except in full, without written permission of Innocal. The results stated in this report relate only to the items tested or calibrated. Measurements reported herein are traceable to SI units via national standards maintained by NIST and were performed in compliance with MIL-STD 45652A, ANSI/NCSL Z540-1:1994, 10 CFR 90, Appendix B, ISO 9002:94, and ISO 17025:2017. Conformance based on Simple Acceptance as a Decision Rule. The estimated measurement uncertainty (EMU), if reported on this certificate, is being reported at a confidence level of 95% or K=2 unless otherwise noted in the remarks section.



NIST Traceable Calibration Report

REPORT NUMBER

1700596

Reference Number: 116455700

PO Number: NBROWN061022

Advanced Scale Inc

13 Delta Dr
 Ste 6
 Londonderry, NH 03053 United States

Manufacturer: Traceable
Model Number: 4096
Description: Humidity, Temp Meter 25 to 95%RH 0-50°C
Asset Number: CP373853
Serial Number: 200074365
Procedure: DS Control Company 4096
Remarks:

Calibration Date: 06/15/2022
Calibration Due Date: 06/15/2024
Condition As Found: In Tolerance
Condition As Left: In Tolerance, No adjustment

NIST-traceable calibration performed on the unit referenced above in accordance with customer requirements, published specifications and the lab's standard operating procedures. No adjustments were made to the unit.

Standards Used

Standard ID	Manufacturer	Model Number	Description	Cal Date	Due Date
CP144802	Rotronic Instrument Corp	HC2A-S	Humidity, HygroClip2	6/21/2021	6/30/2022

Calibration Data

Function Tested	Nominal / Reference Value	Measured Value	OOT	Calibration Tolerance <i>g: = Guard Banding Applied</i>	TUR	EMU
Relative Humidity	50.0 %RH					
As Found & As Left	50.0	49		46.0 to 54.0 %RH	4.9:1	± 0.82 %RH
Temperature	25.0 °C					
As Found & As Left	25.0	24		24.0 to 26.0 °C	8.5:1	± 0.12 °C

Temperature: 26 °C
Humidity: 65 %RH
Rpt. No.: 1700596

Calibration Performed By:				Quality Reviewer:	
Elaine Yonkus	Metrologist			Szplit, Tony	6/15/2022
Name	ID#	Title	Phone	Name	Date

This report may not be reproduced, stored in a retrieval system, or transmitted in any form or by any means, without written permission of Innocal. The results stated in this report relate only to the items tested or calibrated. Measurements reported herein are traceable to SI units via national standards maintained by NIST and were performed in compliance with MIL-STD-45652A, ANSI/NCSL Z540-1:1994, IEC/FR50, Appendix B, ISO 9002:94 and ISO 17025:2017. Conformance based on Simple Acceptance as a Decision Rule. The estimated measurement uncertainty (EMU) reported in this certificate is being reported at a confidence level of 95% or K=2 unless otherwise noted in the remarks section.

625 East Bunker Court
 Vernon Hills, Illinois 60061
 PH: 866-466-6225
 Fax: 847-327-2993
 www.innocalsolutions.com

NIST Traceable
Calibration Report

REPORT NUMBER
1700592

Reference Number: 116455700
 PO Number: NBROWN061022

Advanced Scale Inc

13 Delta Dr
 Ste 6
 Londonderry, NH 03053 United States

Manufacturer:	Traceable	Calibration Date:	06/16/2022
Model Number:	4096	Calibration Due Date:	06/16/2024
Description:	Humidity, Temp Meter 25 to 95%RH 0-50°C	Condition As Found:	In Tolerance
Asset Number:	CP373849	Condition As Left:	In Tolerance, No adjustment
Serial Number:	200074715		
Procedure:	DS Control Company 4096		
Remarks:	NIST-traceable calibration performed on the unit referenced above in accordance with customer requirements, published specifications and the lab's standard operating procedures. No adjustments were made to the unit.		

Standards Used

Standard ID	Manufacturer	Model Number	Description	Cal Date	Due Date
CP144802	Rotronic Instrument Corp	HC2A-S	Humidity, HygroClip2	6/21/2021	6/30/2022

Calibration Data

Function Tested	Nominal / Reference Value	Measured Value	OOT	Calibration Tolerance <i>g: = Guard Banding Applied</i>	TUR	EMU
Relative Humidity	50.0 %RH					
As Found & As Left	50.0	49		46.0 to 54.0 %RH	4.9:1	± 0.82 %RH
Temperature	25.0 °C					
As Found & As Left	25.0	25		24.0 to 26.0 °C	8.5:1	± 0.12 °C

Temperature: 25 °C
 Humidity: 67 %RH
 Rpt. No.: 1700592

Calibration Performed By:				Quality Reviewer:	
Elaine Yonkus	Metrologist			Szplit, Tony	6/16/2022
<small>Name</small>	<small>ID #</small>	<small>Title</small>	<small>Phone</small>	<small>Name</small>	<small>Date</small>

This report may not be reproduced, copied or filed without written permission of Innocal. The results stated in this report relate only to the items tested or calibrated. Measurements reported herein are traceable to SI units via national standards maintained by NIST and were performed in compliance with MIL-STD-45662A, ANSI/NCSL Z540-1:1994, ISO/IEC 17025:2005, ISO 9002:04 and ISO 17025:2017. Conformance based on Simple Acceptance as a Decision Rule. The estimated measurement uncertainty (EMU) reported in this certificate is being reported at a confidence level of 95% or K=2 unless otherwise noted in the remarks section.



625 East Bunker Court
 Vernon Hills, Illinois 60061
 PH: 866-466-6225
 Fax: 847-327-2993
 www.innocalsolutions.com

NIST Traceable
Calibration Report

REPORT NUMBER
1700593

Reference Number: 116455700
 PO Number: NBROWN061022

Advanced Scale Inc

13 Delta Dr
 Ste 6
 Londonderry, NH 03053 United States

Manufacturer:	Traceable	Calibration Date:	06/16/2022
Model Number:	4096	Calibration Due Date:	06/16/2024
Description:	Humidity, Temp Meter 25 to 95%RH 0-50°C	Condition As Found:	In Tolerance
Asset Number:	CP373850	Condition As Left:	In Tolerance, No adjustment
Serial Number:	51147791		
Procedure:	DS Control Company 4096		
Remarks:	NIST-traceable calibration performed on the unit referenced above in accordance with customer requirements, published specifications and the lab's standard operating procedures. No adjustments were made to the unit.		

Standards Used

Standard ID	Manufacturer	Model Number	Description	Cal Date	Due Date
CP144802	Rotronic Instrument Corp	HC2A-S	Humidity, HygroClip2	6/21/2021	6/30/2022

Calibration Data

Function Tested	Nominal / Reference Value	Measured Value	OOT	Calibration Tolerance <i>g: = Guard Banding Applied</i>	TUR	EMU
Relative Humidity	50.0 %RH					
As Found & As Left	50.0	48		46.0 to 54.0 %RH	4.9:1	± 0.82 %RH
Temperature	25.0 °C					
As Found & As Left	25.0	26		24.0 to 26.0 °C	8.5:1	± 0.12 °C

Temperature: 25 °C
 Humidity: 67 %RH
 Rpt. No.: 1700593

Calibration Performed By:				Quality Reviewer:	
Elaine Yonkus		Metrologist		Szplit, Tony	6/16/2022
<small>Name</small>	<small>ID #</small>	<small>Title</small>	<small>Phone</small>	<small>Name</small>	<small>Date</small>

This report may not be reproduced, except in full, without written permission of Innocal. The results stated in this report relate only to the items tested or calibrated. Measurements reported herein are traceable to SI units via national standards maintained by NIST and were performed in compliance with MIL-STD-45652A, ANSI/NCSL Z540-1:1994, 19CFR50, Appendix B, ISO 9002:94 and ISO 17025:2017. Conformance based on Simple Acceptance as a Decision Rule. The estimated measurement uncertainty (EMU), if reported on this certificate, is being reported at a confidence level of 95% or K=2 unless otherwise noted in the remarks section.

625 East Bunker Court
 Vernon Hills, Illinois 60061
 PH: 866-466-6225
 Fax: 847-327-2993
 www.innocalsolutions.com

NIST Traceable
Calibration Report

REPORT NUMBER
1700597

Reference Number: 116455700
 PO Number: NBROWN061022

Advanced Scale Inc

13 Delta Dr
 Ste 6
 Londonderry, NH 03053 United States

Manufacturer:	Traceable	Calibration Date	06/15/2022
Model Number:	4096	Calibration Due Date	06/15/2024
Description:	Humidity, Temp Meter 25 to 95%RH 0-50°C	Condition As Found:	In Tolerance
Asset Number:	CP373854	Condition As Left:	In Tolerance, No adjustment
Serial Number:	200074706		
Procedure:	DS Control Company 4096		
Remarks:	NIST-traceable calibration performed on the unit referenced above in accordance with customer requirements, published specifications and the lab's standard operating procedures. No adjustments were made to the unit.		

Standards Used

Standard ID	Manufacturer	Model Number	Description	Cal Date	Due Date
CP144802	Rotronic Instrument Corp	HC2A-S	Humidity, HygroClip2	6/21/2021	6/30/2022

Calibration Data

Function Tested	Nominal / Reference Value	Measured Value	OOT	Calibration Tolerance <i>g: = Guard Banding Applied</i>	TUR	EMU
Relative Humidity	50.0 %RH					
As Found & As Left	50.0	49		46.0 to 54.0 %RH	4.9:1	± 0.82 %RH
Temperature	25.0 °C					
As Found & As Left	25.0	24		24.0 to 26.0 °C	8.5:1	± 0.12 °C

Temperature: 26 °C
 Humidity: 65 %RH
 Rpt. No.: 1700597

Calibration Performed By:				Quality Reviewer:	
Elaine Yonkus		Metrologist		Szplit, Tony	6/15/2022
<small>Name</small>	<small>ID #</small>	<small>Title</small>	<small>Phone</small>	<small>Name</small>	<small>Date</small>

This report may not be reproduced, except in full, without written permission of Innocal. The results stated in this report relate only to the items tested or calibrated. Measurements reported herein are traceable to SI units via national standards maintained by NIST and were performed in compliance with MIL-STD-45652A, ANSI/NCSL Z540-1:1994, 19 CFR 50, Appendix B, ISO 9002:94 and ISO 17025:2017. Conformance based on Simple Acceptance as a Decision Rule. The estimator measurement uncertainty (EMU), if reported in this certificate, is being reported at a confidence level of 95% or K=2 unless otherwise noted in the remarks section.

625 East Bunker Court
 Vernon Hills, Illinois 60061
 PH: 866-466-6225
 Fax: 847-327-2993
 www.innocalsolutions.com

NIST Traceable
Calibration Report

REPORT NUMBER
1700594

Reference Number: 116455700
 PO Number: NBROWN061022

Advanced Scale Inc
 13 Delta Dr
 Ste 6
 Londonderry, NH 03053 United States

Manufacturer:	Traceable	Calibration Date	06/16/2022
Model Number:	4096	Calibration Due Date	06/16/2024
Description:	Humidity, Temp Meter 25 to 95%RH 0-50°C	Condition As Found:	In Tolerance
Asset Number:	CP373851	Condition As Left:	In Tolerance, No adjustment
Serial Number:	200074366		
Procedure:	DS Control Company 4096		
Remarks:	NIST-traceable calibration performed on the unit referenced above in accordance with customer requirements, published specifications and the lab's standard operating procedures. No adjustments were made to the unit.		

Standards Used

Standard ID	Manufacturer	Model Number	Description	Cal Date	Due Date
CP144802	Rotronic Instrument Corp	HC2A-S	Humidity, HygroClip2	6/21/2021	6/30/2022

Calibration Data

Function Tested	Nominal / Reference Value	Measured Value	OOT	Calibration Tolerance <i>g: = Guard Banding Applied</i>	TUR	EMU
Relative Humidity	50.0 %RH					
As Found & As Left		50.0	48	46.0 to 54.0 %RH	4.9:1	± 0.82 %RH
Temperature	25.0 °C					
As Found & As Left		25.0	25	24.0 to 26.0 °C	8.5:1	± 0.12 °C

Temperature: 25 °C
 Humidity: 67 %RH
 Rpt. No.: 1700594

Calibration Performed By:				Quality Reviewer:	
Elaine Yonkus		Metrologist		Szplit, Tony	6/16/2022
Name	ID #	Title	Phone	Name	Date

This report may not be reproduced, except in full, without written permission of Innocal. The results stated in this report relate only to the items tested or calibrated. Measurements reported herein are traceable to SI units via national standards maintained by NIST and were performed in compliance with MIL-STD-45662A, ANSI/NCSL Z540-1:1994, IEC/ISO Appendix B, ISO 9002:94, and ISO 17025:2017. Conformance based on Simple Acceptance as a Decision Rule. The estimated measurement uncertainty (EMU), if reported in this certificate, is being reported at a confidence level of 95% or K=2 unless otherwise noted in the remarks section.

