

Agency of Agriculture Food & Markets
Consumer Protection Section
116 State Street
Montpelier, VT 05620
agriculture.vermont.gov

Vermont Agriculture & Environmental Laboratory
Vermont Metrology Laboratory
163 Admin Drive
Randolph Center, VT 05061
802-828-2426

Vermont Metrology Laboratory Calibration Certificate

Submitted by:

Advanced Scale
13 Delta Drive Unit 6
Londonderry, NH 03053-2372
603-626-0242

Calibration Date: March 30, 2026
Date Received: March 27, 2026
Calibration Due: Not Specified
Manufacturer: Various
Serial Number: See Chart
Material: Cast Steel
Number of Pieces: 47 - 50 lb, 10 - 25 lb
Description: Field Standards

The mass standard(s) described herein have been compared to the standards of the State of Vermont by NISTIR 6969, SOP 8 (2019), and have been found at time of calibration, or been adjusted, to meet the maximum permissible errors in ASTM E617-23 Standard Specification for Laboratory Weights and Precision Mass Standards. These mass standard(s) were found to have a conventional mass correction at the time of calibration as indicated in the following tabulation. Mass standards are considered within the MPE when the absolute value of the conventional mass correction plus the uncertainty is less than or equal to the specified MPE. Standards received with a conventional mass outside the MPE show a value in the "before adjustment" column.

Standards of the State of Vermont are metrologically traceable to the International System of Units (SI) and the National Institute of Standards and Technology (NIST). The Vermont Laboratory is recognized by NIST under the Laboratory Metrology Program at Mass Echelon III. SI conversion - 1 lb is equal to 0.453 592 37 kg

The uncertainties shown are expressed as the sum of the following sources; (1) Type A, random uncertainties determined by the standard deviation of the measurement process, and (2) Type B, systematic uncertainties relative to the reference standard and procedure used. Type A and Type B uncertainties are combined by the root sum squared method and multiplied by a coverage factor (k) for an approximate 95 % confidence interval.

Technician: Dolan, Kuehne
Mass Comparator(s): MT XP64003L

Environmental Conditions During Calibration

Temperature: 21.6 °C to 21.7 °C
Relative Humidity: 42.4 % to 44.8 %
Barometric Pressure: 763 mmHg to 763 mmHg

Scott Dolan, Weights & Measures Specialist



*The State of Vermont is an Equal Opportunity/Affirmative Action Employer and Provider
Services provided without regard to race, color, national origin, age or disability*

Nominal & Marking	Conventional Mass Correction Before Adjustment	Conventional Mass Correction As Left	Uncertainty	ASTM Class 6 MPE	Units	k Factor
50 lb 200		-1040	110	2300	mg	2.01
50 lb 201		-1920	110	2300	mg	2.01
50 lb 202		-1260	110	2300	mg	2.01
50 lb 203		-1320	110	2300	mg	2.01
50 lb 204		-400	110	2300	mg	2.01
50 lb 205		-970	110	2300	mg	2.01
50 lb 206	-2150	350	110	2300	mg	2.01
50 lb 207		-1950	110	2300	mg	2.01
50 lb 208		-1860	110	2300	mg	2.01
50 lb 209		-490	110	2300	mg	2.01
50 lb 210		-1600	110	2300	mg	2.01
50 lb 211		-70	110	2300	mg	2.01
50 lb 212	-2040	120	110	2300	mg	2.01
50 lb 213		-560	110	2300	mg	2.01
50 lb 214		210	110	2300	mg	2.01
50 lb 215		-500	110	2300	mg	2.01
50 lb 216		110	110	2300	mg	2.01
50 lb 217		-1860	110	2300	mg	2.01
50 lb 218		-730	110	2300	mg	2.01
50 lb 219		-1850	110	2300	mg	2.01
50 lb 400		-1110	110	2300	mg	2.01
50 lb 401		-340	110	2300	mg	2.01
50 lb 402		200	110	2300	mg	2.01
50 lb 403		-170	110	2300	mg	2.01
50 lb 404		-1710	110	2300	mg	2.01
50 lb 405		-330	110	2300	mg	2.01
50 lb 406		20	110	2300	mg	2.01
50 lb 407		220	110	2300	mg	2.01
50 lb 408		-110	110	2300	mg	2.01
50 lb 409		-1560	110	2300	mg	2.01
50 lb 410		-460	110	2300	mg	2.01
50 lb 411		-670	110	2300	mg	2.01
50 lb 412		-1120	110	2300	mg	2.01
50 lb 413		-1500	110	2300	mg	2.01
50 lb 414		-310	110	2300	mg	2.01
50 lb 415		820	110	2300	mg	2.01
50 lb 416		-290	110	2300	mg	2.01
50 lb 417		-560	110	2300	mg	2.01
50 lb 418		-610	110	2300	mg	2.01
50 lb 419		-580	110	2300	mg	2.01
50 lb 420		-1180	110	2300	mg	2.01
50 lb 421		-1010	110	2300	mg	2.01
50 lb 422		-980	110	2300	mg	2.01
50 lb 423		230	110	2300	mg	2.01
50 lb 424		-1120	110	2300	mg	2.01
50 lb 425		-1600	110	2300	mg	2.01
50 lb 426		-1180	110	2300	mg	2.01

Nominal & Marking	Conventional Mass Correction Before Adjustment	Conventional Mass Correction As Left	Uncertainty	ASTM Class 6 MPE	Units	k Factor
25 lb 460		-279	48	1100	mg	2.01
25 lb 461		221	48	1100	mg	2.01
25 lb 462		-614	48	1100	mg	2.01
25 lb 463		61	48	1100	mg	2.01
25 lb 464		-509	48	1100	mg	2.01
25 lb 465		-334	48	1100	mg	2.01
25 lb 466		-619	48	1100	mg	2.01
25 lb 467		-524	48	1100	mg	2.01
25 lb 468		-354	48	1100	mg	2.01
25 lb 469		-474	48	1100	mg	2.01

MPE: Maximum Permissible Error

In addition to meeting ASTM E617-23 Class 6 MPE, all standard also meet NIST Class F Tolerance requirements.

The following weights were adjusted: 206, 212



End of Certificate

*The State of Vermont is an Equal Opportunity/Affirmative Action Employer and Provider
Services provided without regard to race, color, national origin, age or disability*